

Putting New Batteries in the CADD Solis Pump

Follow these steps to change the batteries:

1. Gather these supplies:
 - CADD Solis pump
 - Four AA batteries. (**Only use new batteries. Mixing new and used batteries may cause the low battery alarm not to work.**)
2. Change your batteries every _____.
3. Wash your hands well for 20 seconds. Use liquid soap or hand sanitizer.
4. If you do not have an infusion currently running, skip to step 8.
5. If an infusion is running, stop the pump by pressing the **Stop/Start** button on the front of the pump.
6. When you see, “**Stop pump?**”, select Yes.
7. Press and hold the power switch. This will turn the pump off.
8. To open the battery door, turn the indented notch on the door counterclockwise (to the left). You may use your fingers or a coin.
9. Take out the four old batteries.

10. While holding the pump at an angle, place the four new AA batteries in the pump.
 - Start at the bottom and work your way up.
 - Match the (+) and (-) markings on the new batteries with the markings on the pump door.



Put in new batteries from the bottom up. Match up the (+) and (-) markings.

11. To close the battery door, turn the indented notch clockwise (to the right). You may use your fingers or a coin.
12. Press and hold the power switch to turn the pump on.
 - The pump will run start-up tests, then beep.
 - The main screen will appear.
 - The status bar at the top of the pump will read, “Stopped” and an amber light will blink.

13. Follow the directions you received if you need to start a new infusion.
14. Press **Stop/Start** button to resume the infusion.
15. The screen will read, “Finish interrupted infusion or start from beginning with new bag?” Select Finish.
16. “**Start pump?**” will appear on the screen. Select **Yes**.
17. Status bar will read, “Running,” and a green light will blink. Your medicine is flowing.

Caution: If you put the battery in backwards, the pump will **not** power up. Check the batteries to make sure to match the (+) and (-) markings.

If you have any problems with your infusion, please call us at 612-672-2233 or 800-642-8845.

Did this handout help?

bit.ly/MHF526100

